



## description

Rad-Sure™ blood irradiation indicators provide positive, visual verification of irradiation at the minimum specified dose. Manufactured from Gafchromic™ film, the world's highest resolution dosimeter, Rad-Sure™ has been the standard for blood irradiation indicators for over 25 years. Before a blood product and its attached indicator are irradiated, the indicator reads "NOT IRRADIATED." After the blood product and its attached indicator are irradiated, the word "NOT" is obscured and the indicator reads "IRRADIATED." Rad-Sure™ gamma irradiation indicators should be used with irradiators that have cesium-137 or cobalt-60 radiation sources or other sources producing radiation of equal or greater energy.




- Indicators can now be stored at room temperature
- not subjective the product has been properly irradiated when the film turns black and "NOT" is completely obscured
- color-blind friendly no need to match colors, just read the indicator to know the blood product has been irradiated
- made with accurate and reliable film used in radiation oncology centers around the world
- excellent adhesion to blood bags
- o easy to use: just peel, stick, irradiate and read
- ISBT 128 bar-coded lot numbers
- o minimum dose of 15 Gy or 25 Gy available
- available with no blank fields (NBF) for operator's initials and date



## product list

product	productcode
rad-sure™ 15 Gy blood irradiation indicators	831377
rad-sure™ 25 Gy blood irradiation indicators	831379
rad-sure™ 25 Gy NBF blood irradiation indicators	831380

## Examples



Rad-Sure™ gamma irradiation indicators before and after irradiation







Domestic Orders and Inquiries AdvancedOrders@ashland.com

International Orders and Inquiries
AdvancedMaterialsOrders@ashland.com

## ashland.com/contact

®Registered trademark, Ashland or its subsidiaries, registered in various countries

™Trademark, Ashland or its subsidiaries, registered in various countries

©2018, Ashland / PHA18-108

The information contained in this brochure and the various products described are intended for use only by persons having technical skill and at their own discretion and risk after they have performed necessary technical investigations, tests and evaluations of the products and their uses. Certain end uses of these products may be regulated pursuant to rules or regulations governing medical devices, drug uses, or pesticidal or antimicrobial uses. It is the end user's responsibility to determine the applicability of such regulations to its products.

All statements, information, and data presented herein are believed to be accurate and reliable, but are not to be taken as a guarantee offitness for a particular purpose, or representation, express or implied, for which seller assumes legal responsibility. No freedom to use any patent owned by Ashland, its subsidiaries, or its suppliers is to be inferred.

