



description

Rad-Sure™ XR blood irradiation indicators provide positive, visual verification of irradiation at the minimum specified dose. Manufactured from Gafchromic™ film, the world's highest resolution dosimeter, Rad-Sure™ XR has been the standard for blood irradiation indicators for over 25 years. Before a blood product and its attached indicator are irradiated, the indicator reads "NOT IRRADIATED." After the blood product and its attached indicator are irradiated, the word "NOT" is obscured and the indicator reads "IRRADIATED." They should be used with X-ray irradiators that utilize X-rays generated from 160kVp sources that are filtered through 0.38 mm of copper or 150kVp sources that are filtered through 1 mm of aluminum.



- Indicators can now be stored at room temperature
- o not subjective the product has been properly irradiated when the film turns black and "NOT" is completely obscured
- o color-blind friendly no need to match colors, just read the indicator to know the blood product has been irradiated
- o made with accurate and reliable film used in radiation oncology centers around the world
- excellent adhesion to blood bags
- o easy to use: just peel, stick, irradiate and read
- ISBT 128 bar-coded lot numbers
- o minimum dose of 15 Gy or 25 Gy available
- available with no blank fields (NBF) for operator's initials



product list

product	productcode
rad-sure™ XR 15 Gy blood irradiation indicators	831396
rad-sure™ XR 25 Gy blood irradiation indicators	831398
$\textbf{rad-sure}^{\intercal M} \textbf{ XR 25 Gy NBF} \textbf{ blood irradiation indicators}$	867307

examples



Rad-Sure™ XR irradiation indicators before and after irradiation





IRRADIATED

Ехр. ххххх

Ехр. ххххх



Domestic Orders and Inquiries AdvancedOrders@ashland.com

International Orders and Inquiries
AdvancedMaterialsOrders@ashland.com

ashland.com/contact

- ® Registered trademark, Ashland or its subsidiaries, registered in various countries
- ™ Trademark, Ashland or its subsidiaries, registered in various countries
- © 2018, Ashland / PHA18-108

The information contained in this brochure and the various products described are intended for use only by persons having technical skill and at their own discretion and risk after they have performed necessarytechnical investigations, tests and evaluations of the products and their uses. Certain enduses of these products may be regulated pursuant to rules or regulations governing medical devices, drug uses, or pesticidal or antimicrobial uses. It is the end user's responsibility to determine the applicability of such regulations to its products.

All statements, information, and data presented herein are believed to be accurate and reliable, but are not to be taken as a guarantee offitness for a particular purpose, or representation, express or implied, for which seller assumes legal responsibility. No freedom to use any patent owned by Ashland, its subsidiaries, or its suppliers is to be inferred.

